

United Kingdom Accreditation Service

ACCREDITATION CERTIFICATE



CALIBRATION LABORATORY
No. 0602

Priorclave Ltd

is accredited in accordance with the recognised International Standard ISO/IEC 17025:2005
General Requirements for the competence of testing and calibration laboratories.

This accreditation demonstrates technical competence for a defined scope as detailed in and at the locations specified in the schedule to this certificate, and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated 18 June 2005).

The schedule to this certificate is an essential accreditation document and from time to time may be revised and reissued by the United Kingdom Accreditation Service. The most recent issue of the schedule of accreditation, which bears the same accreditation number as this certificate, is available from the UKAS website www.ukas.org.

This accreditation is subject to continuing conformity with United Kingdom Accreditation Service requirements. The absence of a schedule on the UKAS website indicates that the accreditation is no longer in force.

A handwritten signature in blue ink, appearing to read 'C. Stanger', is written over a horizontal line.

Accreditation Manager, United Kingdom Accreditation Service

Initial Accreditation date
11 June 1998

This certificate issued on
06 April 2006


The Department of Trade and Industry (DTI) has entered into a memorandum of understanding with the United Kingdom Accreditation Service (UKAS) through which UKAS is recognised as the national body responsible for assessing and accrediting the competence of organisations in the fields of calibration, testing, inspection and certification of systems, products and persons

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0602 Accredited to ISO/IEC 17025:2005	Priorclave Ltd	
	Issue No: 007 Issue date: 05 April 2006	
	129/131 Nathan Way West Thamesmead Business Park Woolwich London SE28 0AB	Contact: Mr P R Melling Tel: +44 (0)20-8316 6620 Fax: +44 (0)20-8855 0616 E-Mail: petermelling@priorclave.co.uk Website: www.priorclave.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above*:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of Priorclave)	Temperature (Sterilizers etc, and Thermal Products)	S



0602
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Priorclave Ltd
Issue No: 007 Issue date: 05 April 2006

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, refrigerators, freezers, environmental cabinets, chambers, and rooms (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	-70 °C to 200 °C	0.5 °C*	A list of individual approved engineers is held by the laboratory, and by UKAS Single and multipoint monitoring probes Time dependent temperature profiling	S
TIME INTERVAL Timers	10 s to 2 hrs	2 s	* plus resolution of indicator/recorder	S
END				